



Breast Tomosynthesis

Breast tomosynthesis is an advanced form of mammography, a specific type of breast imaging that uses low-dose x-rays to detect cancer early – before women experience symptoms – when it is most treatable. Breast tomosynthesis is not yet available in all imaging facilities.

Tell your doctor about any breast symptoms or problems, prior surgeries, hormone use, whether you have a family or personal history of breast cancer, and if there's a possibility you are pregnant. If possible, obtain copies of your prior mammograms and make them



available to your radiologist on the day of your exam. Leave jewelry at home and wear loose, comfortable clothing. You may be asked to wear a gown. Don't wear deodorant, talcum powder or lotion under your arms or on your breasts as these may appear on the mammogram and interfere with correct diagnosis.

What is breast tomosynthesis?

Breast tomosynthesis, also called three-dimensional (3-D) mammography and digital breast tomosynthesis (DBT), is an advanced form of breast imaging, or mammography, that uses a low-dose x-ray system and computer reconstructions to create three-dimensional images of the breasts. Breast tomosynthesis aids in the early detection and diagnosis of breast disease.

An x-ray (radiograph) is a noninvasive medical test that helps physicians diagnose and treat medical conditions. Imaging with x-rays involves exposing a part of the body to a small dose of ionizing radiation to produce pictures of the inside of the body. X-rays are the oldest and most frequently used form of medical imaging.

While mammography is the best screening tool for breast cancer available today, it does not detect all breast cancers. Breast tomosynthesis overcomes some of the limitations of standard mammography, but it is not yet available in all imaging facilities.

A conventional x-ray examination of the breast, called a mammogram, is two-dimensional: two x-ray images are taken of the breast, from top-to-bottom and from angled side-to-side, while the breast is

compressed between a clear plastic paddle and an imaging detector. Although compression is necessary to obtain breast images, it may cause overlapping of the breast tissue in which abnormal tissue can be hidden and superimposed normal tissue can appear abnormal.

In breast tomosynthesis, the x-ray tube moves in an arc over the compressed breast capturing multiple images of each breast from different angles. These digital images are then reconstructed or “synthesized” into a set of three-dimensional images by a computer. These three dimensional image sets help minimize the tissue overlap that can hide cancers or make it difficult to distinguish normal overlapping breast tissue from tumors.

What are some common uses of the procedure?

Breast tomosynthesis can be used to perform screening mammography, a test to detect early breast cancer in women experiencing no symptoms. Screening mammography with breast tomosynthesis has been shown to improve accuracy and reduce false positive rates in women of all breast densities. The largest improvements with tomosynthesis are seen in women with dense breast tissue, but women with non-dense breasts also demonstrate significantly better cancer detection and significantly fewer false positive findings when tomosynthesis is used. See the [Dense Breasts page](#) for more information about dense breasts.

Breast tomosynthesis may also be used to perform diagnostic mammography to detect and diagnose breast disease in women experiencing symptoms such as a lump, pain, skin dimpling or nipple discharge.

Screening Mammography with Breast Tomosynthesis

Mammography plays a central part in the early detection of breast cancers because it can often show changes in the breast years before a patient or physician can feel them. Current guidelines from the U.S. Department of Health and Human Services (HHS) and the American College of Radiology (ACR) recommend screening mammography every year for women beginning at age 40. Research has shown that annual mammograms lead to early detection of breast cancers when they are most curable and breast-conservation therapies are available. See the [Mammography page](#) for more information about mammography.

The National Cancer Institute (NCI) advises women who have had breast cancer and those who are at increased risk due to a family history of breast cancer to seek expert medical advice about the frequency of screening and whether they should begin screening before age 40. According to American Cancer Society guidelines, most women at high risk should begin screening with MRI and mammography at age 30 and continue for as long they are in good health. See the [Breast Cancer Screening page](#) for information about breast cancer screening.

Diagnostic Mammography with Breast Tomosynthesis

Diagnostic mammography is used to evaluate a patient with abnormal clinical findings—such as a breast lump or nipple discharge—that have been found by a patient or physician. Diagnostic mammography may also be done after an abnormal screening mammogram in order to evaluate the area of concern on the

screening exam.

How should I prepare?

Before scheduling a mammogram, the American Cancer Society (ACS) and other specialty organizations recommend that you discuss any new findings or problems in your breasts with your doctor. In addition, inform your doctor of any prior surgeries, hormone use, and family or personal history of breast cancer.

Do not schedule your screening mammogram for the week before your menstrual period if your breasts are usually tender during this time. The best time for a screening mammogram is one week following your period. Always inform your doctor or x-ray technologist if there is any possibility that you are pregnant.

The ACS also recommends you:

- Do not wear deodorant, talcum powder or lotion under your arms or on your breasts on the day of the exam. These can be confused for calcium deposits on the mammogram.
- Describe any breast symptoms or problems to the technologist performing the exam.
- Obtain your prior mammograms and make them available to the radiologist if they were done at a different location. This is needed for comparison with your current exam and can often be obtained on a CD.
- Ask when your results will be available; do not assume the results are normal if you do not hear from your doctor or the mammography facility.

What does the equipment look like?

A mammography unit is a rectangular box that houses the tube in which x-rays are produced. The unit is used exclusively for x-ray exams of the breast, with special accessories that allow only the breast to be exposed to the x-rays. Attached to the unit is a device that holds and compresses the breast and positions it so images can be obtained at different angles.

Breast tomosynthesis is performed using a digital mammography unit, but not all digital mammography machines are equipped to perform tomosynthesis imaging.

How does the procedure work?

Breast tomosynthesis uses a low-dose x-ray system, electronics and a computer to convert x-ray images of the breast to a three-dimensional image set. Multiple x-ray images of the breast are digitized using systems similar to those found in digital cameras and then transferred to a computer where they are reconstructed or “synthesized” into a three-dimensional image set. In this way, 3-D breast imaging is similar to computed tomography (CT) imaging in which a series of thin high-resolution “slices” are assembled together to create a 3-D reconstruction of the body.

X-rays are a form of radiation like light or radio waves. X-rays pass through most objects, including the body. Once it is carefully aimed at the part of the body being examined, an x-ray machine produces a small burst of radiation that passes through the body, recording an image on photographic film or a special detector.

Different parts of the body absorb the x-rays in varying degrees. Dense bone absorbs much of the radiation while soft tissue, such as muscle, fat and organs, allow more of the x-rays to pass through them. As a result, bones appear white on the x-ray, soft tissue shows up in shades of gray and air appears black.

Until recently, x-ray images were maintained on large film sheets (much like a large photographic negative). Today, most images are digital files that are stored electronically. These stored images are easily accessible and are frequently compared to current x-ray images for diagnosis and disease management.

How is the procedure performed?

Breast tomosynthesis is performed on an outpatient basis.

During this exam, a specially qualified radiologic technologist will position your breast in the mammography unit. Your breast will be placed on a special platform and gradually compressed with a clear plastic paddle. Breast compression is necessary during tomosynthesis imaging in order to:

- Even out the breast thickness so that all of the tissue can be visualized.
- Spread out the tissue so that small abnormalities are less likely to be hidden by overlying breast tissue.
- Allow the use of a lower x-ray dose since a thinner amount of breast tissue is being imaged.
- Hold the breast still in order to minimize blurring of the image caused by motion.

Your breast will remain compressed while the x-ray tube moves from one side of your breast to the other in an arc, capturing multiple images from different angles along the way. In addition to the three-dimensional image set, the breast tomosynthesis exam will also synthesize standard two-dimensional images of the breast for interpretation.

You must hold very still and may be asked to keep from breathing for a few seconds while the x-ray picture is taken to reduce the possibility of a blurred image. The technologist will walk behind a wall or into the next room to activate the x-ray machine.

When the examination is complete, you may be asked to wait until the radiologist determines that all the necessary images have been obtained.

The examination process should take about 30 minutes.

What will I experience during and after the procedure?

You will feel pressure on your breast as it is squeezed by the compression paddle. Some women with

sensitive breasts may experience discomfort. If this is the case, schedule the procedure when your breasts are least tender. Be sure to inform the technologist if pain occurs as compression is increased. If discomfort is significant, less compression will be used. Always remember compression allows better quality mammograms, including breast tomosynthesis exams.

Who interprets the results and how do I get them?

A radiologist, a physician specifically trained to supervise and interpret radiology examinations, will analyze the images and send a signed report to your primary care or referring physician, who will discuss the results with you.

You will also be notified of the results by the mammography facility.

Follow-up examinations may be necessary. Your doctor will explain the exact reason why another exam is requested. Sometimes a follow-up exam is done because a potential abnormality needs further evaluation with additional views or a special imaging technique. A follow-up examination may also be necessary so that any change in a known abnormality can be monitored over time. Follow-up examinations are sometimes the best way to see if treatment is working or if a finding is stable or changed over time.

What are the benefits vs. risks?

Benefits

- Imaging of the breast allows detection of small tumors. When cancers are small, the woman has more treatment options.
- The use of screening mammography increases the detection of small abnormal tissue growths confined to the milk ducts in the breast, called ductal carcinoma in situ (DCIS). These early tumors rarely harm patients if they are removed at this stage and mammography is an excellent way to detect these tumors. It is also useful for detecting all types of breast cancer, including invasive ductal and invasive lobular cancer.
- No radiation remains in a patient's body after an x-ray examination.
- X-rays usually have no side effects in the typical diagnostic range for this exam.
- Large population studies have shown that screening with breast tomosynthesis results in improved breast cancer detection rates and fewer “call-backs,” instances where women are called back from screening for additional testing because of a potentially abnormal finding.
- Breast tomosynthesis may also result in:
 - earlier detection of small breast cancers that may be hidden on a conventional mammogram
 - greater accuracy in pinpointing the size, shape and location of breast abnormalities
 - fewer unnecessary biopsies or additional tests
 - greater likelihood of detecting multiple breast tumors
 - clearer images of abnormalities within dense breast tissue

Risks

- There is always a slight chance of cancer from excessive exposure to radiation. However, the benefit of an accurate diagnosis far outweighs the risk.
- The effective radiation dose for this procedure varies. See the Safety page for more information about radiation dose.
- **False Positive Mammograms.** Five percent to 15 percent of screening mammograms require more testing such as additional mammograms or ultrasound. Most of these tests turn out to be normal. If there is an abnormal finding, a follow-up or biopsy may have to be performed. Most biopsies confirm that no cancer is present. The likelihood of false positive findings and negative biopsies is reduced when tomosynthesis imaging is used for screening and diagnostic mammographic examinations.
- Women should always inform their physician or x-ray technologist if there is any possibility that they are pregnant. See the Safety page for more information about pregnancy and x-rays.

A Word About Minimizing Radiation Exposure

Special care is taken during x-ray examinations to use the lowest radiation dose possible while producing the best images for evaluation. National and international radiology protection organizations continually review and update the technique standards used by radiology professionals.

Modern x-ray systems have very controlled x-ray beams and dose control methods to minimize stray (scatter) radiation. This ensures that those parts of a patient's body not being imaged receive minimal radiation exposure.

Although the radiation dose for breast tomosynthesis is slightly higher than the dosage used in standard mammography, it remains within the FDA-approved safe levels for radiation from mammograms. Some systems have doses very similar to conventional mammography.

What are the limitations of breast tomosynthesis?

Interpretations of mammograms can be difficult because a normal breast looks different for each woman. Also, the appearance of an image may be challenging if you have undergone breast surgery. Because some breast cancers are hard to visualize, a radiologist may want to compare the image to views from previous examinations. It is very important to realize that not all breast cancers can be seen on mammography.

Breast implants may also impede accurate mammogram readings because both silicone and saline implants are not transparent on x-rays and can block a clear view of the tissues behind them, especially if the implant has been placed in front of the chest muscles. Experienced technologists and radiologists know how to carefully compress the breasts and move the implants out of the image to improve the view without rupturing the implant.

While mammography is the best screening tool for breast cancer available today, mammograms do not

detect all breast cancers. This is called a false negative result. On the other hand, when a mammogram looks abnormal and no cancer is present, this is called a false-positive result.

Advanced imaging techniques such as breast tomosynthesis are helping to overcome the limitations of conventional mammography. Research continues on breast tomosynthesis and other breast imaging techniques that can contribute to the early detection of breast cancer and improve the accuracy in distinguishing non-cancerous breast conditions from breast cancers.

Disclaimer

This information is copied from the RadiologyInfo Web site (<http://www.radiologyinfo.org>) which is dedicated to providing the highest quality information. To ensure that, each section is reviewed by a physician with expertise in the area presented. All information contained in the Web site is further reviewed by an ACR (American College of Radiology) - RSNA (Radiological Society of North America) committee, comprising physicians with expertise in several radiologic areas.

However, it is not possible to assure that this Web site contains complete, up-to-date information on any particular subject. Therefore, ACR and RSNA make no representations or warranties about the suitability of this information for use for any particular purpose. All information is provided "as is" without express or implied warranty.

Please visit the RadiologyInfo Web site at <http://www.radiologyinfo.org> to view or download the latest information.

Note: Images may be shown for illustrative purposes. Do not attempt to draw conclusions or make diagnoses by comparing these images to other medical images, particularly your own. Only qualified physicians should interpret images; the radiologist is the physician expert trained in medical imaging.

Copyright

This material is copyrighted by either the Radiological Society of North America (RSNA), 820 Jorie Boulevard, Oak Brook, IL 60523-2251 or the American College of Radiology (ACR), 1891 Preston White Drive, Reston, VA 20191-4397. Commercial reproduction or multiple distribution by any traditional or electronically based reproduction/publication method is prohibited.

Copyright © 2017 Radiological Society of North America, Inc.